NUCLEONIX

NEUTRON AREA MONITOR

TYPE:NXG_NAM4_2

Technical Data



FEATURES :

- Microcontroller based design.
- Uses He3 proportional detector.
- 9" dia polyethylene moderator covers the detector.
- Measures neutron flux in CPM & neutron dose rate in uSv/hr.
- Automatic TC selection depending upon the countrate inversely varying with countrate.
- Large area alarm annunciator windows for NORMAL & ACTIVE conditions.
- 4-20mA current loop 0-5V analog signal for remote indication on a I/O connector.
- Ethernet RS 485 port with MODBUS RTU compliant firmware.

Neutron Area Monitor NXG_NAM4_2 manufactured by Nucleonix Systems designed using advanced technologies is primarly meant for detecting neutron flux leakages and measure dose rate in reprocessing plants, atomic power plants accelerator buildings and medical cyclotron facilities. It serves as a neutron zone monitor to indicate flux / countrate or dose rate and gives alarms visual / aural once the preset levels fixed by the user exceeds. Also relays will be activated.

This unit indicates the dose rate digitally on a 4 x 7 segment LED display / TFT display. Each of the annunciator windows for NORMAL and ACTIVE conditions has LED array. Once alarm triggers ACTIVE window starts blinking.

Unit can be programmed/configured using front panel keypad which can be deactivated after completion programming. Configuring the unit namely setting preset level, setting reset mode - AUTO/ MANUAL etc are achieved by this keypad. Unit also performs self-diagnostics for HV failure, pulse processing electronics failure & detectors failure on power up.

Alarm acknowledge and reset pins are provided on the circular I/O connector for remote acknowledge & reset. Ethernet RS485 port is also available optionally.

NUCLEONIX

STANDARD SPECIFICATIONS

1. Radiation detected :

Neutron Flux.

2. Detector :

3. **Detector Type :** He3 neutron Proportional detector.

- 4. Gas pressure : 3800 (torr).
- 5. Cathode Material : SS.
- 6. Opearting Voltage: 1300+/- 50V
- 7. Sensitivity : 38CPS / Nv

8. **Detector assembly :** Surrounded by 9" dia polyethylene moderator.

9. Energy Response : Energy dependent.

10. **Range of measurement:** (0-10)mSv/hr.

11. Gamma Energy Response: <3 CPM(Avg) for 10R/hr Co 60.

12. **Time constant :** Inversely varying with countrate.

13. **Dispaly** : TFT / 6 digit 7 segment LED display for display of neutron flux 16x2 LCD display for visualization of preset alarm & other parameters.

14. **Time Constant** :16 sec to 0.5 sec inversely varying with countrate.

15. **Alarm range :** Available throughout the range

16. Alarm System :

a.Alarm will be indicated by

Red (LED) flashing large area window display and Loud audio alarm.

b. Distinctive EHT / Detector failure alarm.

c.Large area instrument "ON" indication.

d. Provision for display of alarm set point.

e.Alarm set – point will be adjustable continuously.

f. The instrument will have alarm acknowledgement and Reset switches.

g. Provision for remote alarm acknowledge and reset.

h. Alarm annunciation scheme: as tabulated below

17. Monitor Enclosure

Vapour-tight, rugged & elegant. The door shall be lockable with knob. The instrument can be offered with compliance to required IP standards.

18. Remote/External console

a.4-20 mA full-scale output. Output circuitry shall be able to drive 200 mtrs. of twisted pair of wires.
b. Two sets of potential free contacts of Alarm relay (Change over).
Contact rating 3 Amps at 250 VAC.
c.Remote alarm acknowledgement and reset signals for the field instruments.

19. Self-Diagnostics

The monitor will have built-in selfdiagnostics. On being powered it will perform tests to ensure that all components and sub systems are functioning properly. It will check for the High voltage supply, Detector and pulse processing electronics.

20. Power

230 VAC +/-10%, 50 Hz, single-phase supply. Line filter with transient suppressor is built-in.

21. Environment

The instrument will be able to withstand temperature upto 50° C and relative humidity upto 90% in radiation areas.Instrument is ISO9000 part II & part IV.

OPTIONAL (AT EXTRA COST)

1.Computer Interface

The Monitor will have a Ethernet / RS 485 serial Communication port for interfacing with a PC. The PC and the monitor will operate in a host slave configuration in a multi drop network through this interface.

The PC, as the host will give commands and sens queries. The monitor will carry out the various functions as per the required information in response to the queries.

The firmware of the monitor will be able to send the instrument data like instrument ID, instrument type, input range, display range, alarm current settings,alarm data, readings etc to the Host PC and carry out the setting of different parameters like instrument input range, id,instrument type, display range, alarm settings etc. The user will provide detailed list of the command and response for the Hostslave communication.

The detailed specifications for the interface will be as follows :

Type: Ethernet/RS-485 Multi drop Serial Communication Port,Half-duplex bi-directional communi-cation.

Protocol : MODBUS RTU.

Bit Rate : Baud rate will be 19,200 by default. VB software & configurator will be able to communicate at these baud rates. For operating at other baud rates PC will modify the baud rate settings of the Area Monitor.

Address :User configurable from 0 to 255. Device address be entered is to PC.Address range 0-255 using is Each of the units should have distinct address. This address is stored in

EEPROM and whenever an address command is received from central console, if address matches to this monitor, it will respond to the queries.

Applications : It measures Neutron Dose Rate & gives Alarms Annunciation. Recomended in Nuclear Plants, &

Acclerator facilities. Where neutron radiation levels are to be monitored continuously.

Connector:

9 pin D-type connector.

2. EMI / EMC Compliance :

Unit can be offered with EMI/EMC complaince on request at extra cost.

3. IEC Enclosure Standard :

The instrument can be offered with complaince to required IP standards.